



ecology and environment, inc.

101 YESLER WAY, SEATTLE, WASHINGTON, 98104, TEL. 206/624-9537

International Specialists in the Environment

USEPA SF



1025422

31108
1.9.4.3

MEMORANDUM

DATE: March 22, 1988

TO: Priscilla Anderson, TATM-Toxicologist, E&E, Seattle

FROM: Roger McGinnis, Chemist, E&E, Seattle *Rmm*

SUBJ: Polychlorinated dibenzodioxin/furan Homologue Analysis Data
Quality Assurance Review for Corigliano Landfill

REF: TDD: T10-8707-009
PAN: TWA-0526-RFA

The quality assurance review of two soil samples collected from Corigliano Landfill has been completed. The samples were analyzed for polychlorinated dibenzodioxins/furans (PCDD/PCDF) by TMS Analytical Services of Indianapolis, Indiana.

The samples were numbered:

T7080035 and T7080036.

Data Qualifications

Samples were analyzed by Tandem Mass spectroscopy (MS/MS). While no USEPA protocols have been established for this method data review was performed using modified USEPA Method 8280 criteria.

I Timeliness

No holding time criteria have been established. Sample T7080036 was extracted and analyzed on the day received. Sample T7080035 was analyzed 24 days after receipt.

II PCDD/PCDF Analysis

The laboratory did not specify the chromatographic column used for separation. However, they stated a short, non-polar column was used rather than the standard, longer, polar column.

III Initial Calibration

a) Sensitivity

The signal/noise ratio for internal standards was greater than 10.

31108

b) Relative Response Factors

No criteria have been established for internal standard RRFs. The data summary sheet was missing for run 1 of standard solution 1.

c) Linearity

Internal standard RRFs met the laboratory criteria of less than 10% relative standard deviation (RSD).

d) Ion Ratios

Laboratory criteria for ion ratios were met. Ratios were independent of concentration with less than 10% RSD.

e) Surrogate Standards

No surrogate raw data including areas, ion ratios or relative response factors were supplied by the laboratory.

IV Continuing Calibration

a) Relative Response Factors

Percent difference between initial calibration and continuing calibration RRFs exceeded 10% for a number of dioxin/furan congeners. Tabulated areas for HpCDD/HpCDF do not match raw data for sample T7080035.

<u>Date</u>	<u>Congener</u>	<u>% Difference</u>	<u>QA Limit</u>
8/28/87	TCDD	12.3%	10%
8/28/87	TCDF	24.0%	10%
8/28/87	PCDD	11.0%	10%
8/28/87	PCDF	21.6%	10%
8/28/87	HPCDF	10.5%	10%
8/28/87	OCDF	12.4%	10%
9/21/87	TCDF	14.7%	10%
9/21/87	PCDF	21.6%	10%

The laboratory did not perform a new initial calibration as required by their quality control protocols.

b) Ion Ratios

The percent difference between initial and continuing calibration ion ratios was 14.4% for OCDD. The laboratory did not perform a new initial calibration.

V Laboratory Method Blank

No chlorinated dioxins or furans were detected in the laboratory blank. The laboratory stated ion ratios did not meet criteria for

positive identification though no raw data were provided.

VI Duplicate Sample Analysis

No duplicate analysis was performed.

VII Native Spike Analysis

No native spike analysis was performed by the laboratory.

VIII Field Blank Analysis

No field blank was submitted to the laboratory.

IX Performance Evaluation Sample

No performance evaluation sample was submitted to the laboratory.

X Surrogate Analysis

Reported Surrogate recoveries for several samples were outside laboratory quality control limits. No data were present to verify results.

<u>Sample</u>	<u>Surrogate</u>	<u>% Recovery</u>	<u>QA Limits</u>
Method Blank	Hepta/Octa	31.88	60-140%
T7080035	Hexa/Hepta	54.01	60-140%
T7080035	Hepta/Octa	38.38	60-140%
T7080036	Tetra/penta	196.27	60-140%

XI Sample Analysis

Ion chromatograms of sample T7080035 showed peaks in dioxin/furan windows. The laboratory stated ion ratios did not meet matching criteria though no data were present to verify ratios.

Sample T7080036 contained positively identified dioxin/furan homologues. Concentrations were at or slightly below stated laboratory quantitation limits.

XII Data Assessment

Data copy quality was unacceptable. Many raw data sheets were illegible and poorly labeled.

Sample quantitation limits are considered estimated quantities only (flagged J) since no data are present regarding duplicate, native spike, and performance evaluation samples. In addition, the continuing calibration did not meet criteria.

It should be noted sample T7080036 contained positively identified dioxin/furan homologues at estimated concentrations slightly below estimated laboratory quantitation limits. True concentrations may be higher or lower.



TMS ANALYTICAL SERVICES, INC.

6376 Morenci Trail
Indianapolis, Indiana 46268
317 291-5607

SCHEME MS ANALYSIS REPORT FORM ANALYSIS FOR POLYCHLORINATED DIOXINS AND FURANS

CLIENT: ECOLOGY & ENVIRONMENT
SAMPLE: 77080035

ANALYTE	RESULT	ESTIMATED LIMIT	TODD	ESTIMATED LIMIT
	TOTAL PCDD/PCDF'S	OF DETECTION	RESULT	OF DETECTION
	ng/g	ng/g	2,3,7,8-WINDOW **	2,3,7,8-WINDOW **
=====	=====	=====	=====	=====
TCDD = TETRACHLORODIBENZODIOXIN	NONE DETECTED	J 0.23	NONE DETECTED	J 0.23
TCDF = TETRACHLORODIBENZOFURAN	NONE DETECTED	0.21		
PCDD = PENTACHLORODIBENZODIOXIN	NONE DETECTED	0.20		
PCDF = PENTACHLORODIBENZOFURAN	NONE DETECTED	0.19		
HCDD = HEXACHLORODIBENZODIOXIN	NONE DETECTED	0.76		
HCDF = HEXACHLORODIBENZOFURAN	NONE DETECTED	0.51		
HPCDD = HEPTACHLORODIBENZODIOXIN	NONE DETECTED	1.56		
HPCDF = HEPTACHLORODIBENZOFURAN	NONE DETECTED	1.09		
OCDD = OCTACHLORODIBENZODIOXIN	NONE DETECTED	4.36		
OCDF = OCTACHLORODIBENZOFURAN	NONE DETECTED	3.60		

*** See QA memo
Dioxins / Furans may be
present at low
concentrations

ng/g = parts-per-billion
ng/kg = parts-per-trillion

AN ACCEPTABLE METHOD BLANK WAS RUN ALONG WITH THIS SAMPLE; THE METHOD BLANK SHOWED
NONE-DETECTED RESULTS FOR ALL ANALYTES

LOD = LIMIT OF DETECTION CALCULATED AS EQUIVALENT OF 2.5 X PEAK-TO-VALLEY NOISE PEAK HEIGHT OF IND DAUGHTER ION
COMPARED TO PEAK HEIGHT OF INTERNAL STANDARD.

NOTE: FOR ALL POSITIVE RESULTS THE ION RATIOS FOR DAUGHTER IONS WERE ALL WITHIN 15% OF CALIBRATION AVERAGES

** NOTE: THE 2,3,7,8-TODD WINDOW IS ONLY SEMI-ISOMER SPECIFIC IN THAT A SHORT, NON-POLAR
CAPILLARY COLUMN WAS USED FOR SCREENING PURPOSES. THIS COLUMN SEPARATES THE
THE 2,3,7,8-TODD ISOMER FROM ALL OTHER ISOMERS EXCEPT THE TWO CLOSEST ELUTING ISOMERS.
ACCORDINGLY, THIS NUMBER SHOULD BE REGARDED AS THE MAXIMUM 2,3,7,8-TODD
TO BE EXPECTED UPON ISOMER SPECIFIC ANALYSIS. TO PROPERLY USE THIS
RESULT, IT IS ESSENTIAL TO UNDERSTAND THAT UPON ANALYSIS USING A LONGER
POLAR CAPILLARY COLUMN MEETING EPA CRITERIA FOR ISOMER SPECIFICITY, THAT
ANY VALUE, INCLUDING ZERO, BELOW THIS VALUE MAY BE OBTAINED. SUCH WIDE DIFFERENCES
ARE RELATIVELY RARE, AND SINCE THE ASSUMPTIVE ERR ON THE SIDE OF SAFETY
THE VALUES ARE USEFUL FOR SCREENING PURPOSES. NEGATIVE, OR NONE-DETECTED RESULTS
ARE DEFINITIVE AND NO FURTHER WORK NEED BE DONE EVEN IF ISOMER SPECIFICITY
IS DESIRED.



TMS ANALYTICAL SERVICES, INC.

6576 Morena Trail
Indianapolis, Indiana 46268
317-291-5697

GC/MS/MS ANALYSIS REPORT FORM ANALYSIS FOR POLYCHLORINATED DIOXINS AND FURANS

CLIENT: ECOLOGY & ENVIRONMENT
SAMPLE: T7080036

ANALYTE	RESULT TOTAL PCDD/PCDF'S ng/g	ESTIMATED LIMIT OF DETECTION ng/g	TODD	ESTIMATED LIMIT
			RESULT 2,3,7,8-WINDOW ** ng/g	OF DETECTION 2,3,7,8-WINDOW ** ng/g
TODD = TETRACHLORODIBENZODIOXIN	NONE DETECTED	J 0.38	NONE DETECTED	J 0.38
TODF = TETRACHLORODIBENZOFURAN	NONE DETECTED	0.22		
PCDD = PENTACHLORODIBENZODIOXIN	NONE DETECTED	0.23		
PCDF = PENTACHLORODIBENZOFURAN	NONE DETECTED	0.26		
HxCDD = HEXACHLORODIBENZODIOXIN	NONE DETECTED	0.54		
HxCDF = HEXACHLORODIBENZOFURAN	NONE DETECTED	0.60		
HxCDD = HEPTACHLORODIBENZODIOXIN	NONE DETECTED	0.97		
HxCDF = HEPTACHLORODIBENZOFURAN	NONE DETECTED	1.13		
OCDD = OCTACHLORODIBENZODIOXIN	NONE DETECTED	1.78		
OCDF = OCTACHLORODIBENZOFURAN	NONE DETECTED	1.44		

*** See QA memo
Dioxins / Furans were
positively identified
at low concentrations

ng/g = parts-per-billion
ng/kg = parts-per-trillion

AN ACCEPTABLE METHOD BLANK WAS RUN ALONG WITH THIS SAMPLE; THE METHOD BLANK SHOWED
NONE-DETECTED RESULTS FOR ALL ANALYTES

LLO = LIMIT OF DETECTION CALCULATED AS EQUIVALENT OF 2.5 X PEAK-TO-VALLEY NOISE PEAK HEIGHT OF 2ND DAUGHTER ION
COMPARED TO PEAK HEIGHT OF INTERNAL STANDARD.

NOTE: FOR ALL POSITIVE RESULTS THE ION RATIOS FOR DAUGHTER IONS WERE ALL WITHIN 15% OF CALIBRATION AVERAGES

** NOTE: THE 2,3,7,8-TODD WINDOW IS ONLY SEMI-ISOMER SPECIFIC IN THAT A SHORT, NON-POLAR
CAPILLARY COLUMN WAS USED FOR SCREENING PURPOSES. THIS COLUMN SEPARATES THE
THE 2,3,7,8-TODD ISOMER FROM ALL OTHER ISOMERS EXCEPT THE TWO CLOSEST ELUTING ISOMERS.
ACCORDINGLY, THIS NUMBER SHOULD BE REGARDED AS THE MAXIMUM 2,3,7,8-TODD
TO BE EXPECTED UPON ISOMER SPECIFIC ANALYSIS. TO PROPERLY USE THIS
RESULT, IT IS ESSENTIAL TO UNDERSTAND THAT UPON ANALYSIS USING A LONGER
POLAR CAPILLARY COLUMN MEETING EPA CRITERIA FOR ISOMER SPECIFICITY, THAT
ANY VALUE, INCLUDING ZERO, BELOW THIS VALUE MAY BE OBTAINED. SUCH WIDE DIFFERENCES
ARE RELATIVELY RARE, AND SINCE THE ASSUMPTIONS EFF ON THE SIDE OF SAFETY
THE VALUES ARE USEFUL FOR SCREENING PURPOSES. NEGATIVE, OR NONE-DETECTED RESULTS
ARE DEFINITIVE AND NO FURTHER WORK NEED BE DONE EVEN IF ISOMER SPECIFICITY
IS DESIRED.